

Presented by: Michael Kramer Calibration/Inspection Program Manager Perry Johnson Laboratory Accreditation, Inc. <u>mkramer@pjlabs.com</u> 23-July-2021



This webinar is being recorded and will be available in it's entirely on the Perry Johnson Laboratory Accreditation Website. <u>www.pjlabs.com</u>

Go to the link for recorded webinars.

Duration of webinar is set for one hour.

You can type any questions directly into your webinar box; We will review them at the conclusion of today's session;





#### Update Notification 55 "April 2, 2021"





### *PJLA Policy on Measurement Uncertainty "PL-3"* The requirement to estimate CMC (Calibration and Measurement Capability)applies to calibration organizations only. This policy is based on the requirements outlined in ISO/IEC 17025:2017, ISO 15189:2012, ISO 17034:2016, ISO/IEC 17011:2017 and ILAC P-14:09/2020 and applies only to calibrations or tests for which an accredited result is to be reported.

In accordance with ILAC P14, laboratories will be required to

determine measurement uncertainty in accordance with the GUM "Guide to the Expression of Uncertainty in Measurement"





**Calibration and Measurement Capabilities Calibration and Measurement Capabilities:** Is an effort to express "The smallest uncertainty which an organization can attain when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions".

Accredited calibration laboratories shall not report a smaller measurement uncertainty than the uncertainty described by the CMC for which the laboratory is accredited

e Te	Tools Updat	e_55.pdf	PL-3 Uncertaint	y.pdf Pressu	re Gage Cert ×					(?)	Sig	gn
25			Ť	1 / 2		135% -				O-		1
	Loc/Dept.:	NA/NA			Procedure:	T.O. 33	K6-4-427-1	<b>Rev:</b> 04	/2019		-	
	Technician:	Steven Housi	Housley-Follis Units: psi Range: 200									
	Comments:	NA					Resolution: 5					
	Service Date:	07/15/2020	Temp:	21.4° C	Humidity: 46.2%	ſ	Received Condition	e				
	Due Date:	07/15/2021	Barome	tric Pressure:	N/A		Returned Condition			-		
							Readi	ngs —				
	Test Point Para	meters	UOM	Nominal	Tolera	nce —	Before*	After*	UNC			
	Increasing Pres	sure	psi	0.00	+6.00	-6.00	0.00	0.00	5.8			١.
			psi	40.00	+6.00	-6.00	42.28	42.28	5.8			
			psi	80.00	+6.00	-6.00	80.57	80.57	5.8			
			psi	120.00	+6.00	-6.00	120.15	120.15	5.8			
			psi	160.00	+6.00	-6.00	160.27	160.27	5.8			
			DSI	200.00	+6.00	-6.00	202.89	202.89	5.8			
	Decreasing Pre	ssure	psi	80.00	+6.00	-6.00	81.08	81.08	5.8			
	Percenting metabolis indicate they are out of tolerance.											
	*Bold reading en	and a manual e the	, a.e. 501 01 10									
	*Bold reading en											
	*Bold reading en			Calibra	tion Standards	Used						



Accredited organizations shall identify the contributions to measurement uncertainty. When evaluating measurement uncertainty, all contributions that are of significance, including those arising from sampling, shall be taken into account using appropriate methods of analysis. The organization must estimate the CMC for every measured quantity, instrument or gauge listed in its desired scope of accreditation in accordance with its documented procedure.

The organization shall then prepare an uncertainty budget containing all relevant information related to the identified significant sources of

uncertainty.

Source of Uncertainty	Value Units a <sub>i</sub>		Probability Distribution	Divisor	Sensitivity Coefficient <i>c<sub>i</sub></i>	Standard Uncertainty U <sub>i</sub> (y) (mm)						
Calibration Uncertainty	0.01	mm	Normal (k=2)	2	1	0.005						
Resolution 0.005 mm		mm	Triangular	√6	1	0.002						
Cosine error	3 deg		Rectangular	√3	0.046	0.080						
Temperature	2	С	Rectangular	√3	0.0023	0.003						
Repeatability 0.02		mm	Normal (k=1)	1	1	0.020						
Combined Standard Uncertainty u <sub>c</sub> (y)												
Expanded Uncertainty (k=2, 95% confidence) U												



*Necessary steps in developing an estimate of measurement uncertainty Step 1* <u>*Identify:*</u> *Make a list of all equipment or conditions that diminish the "correctness" of the measurement result* 

Step 2 <u>Quantify:</u> Determine reasonable values for the standard uncertainty of each identified contributor.

Step 3 <u>Combine:</u> Combine all standard uncertainties using the RSS (Root Sum of Squares) method.

$$u_c = \sqrt{u_1^2 + u_2^2 + u_3^2 + u_4^2 + u_5^2 + u_6^2}$$

Step 4 **<u>Expand the uncertainty</u>**. Multiply by the appropriate coverage factor to obtain the expanded uncertainty of the measurement result

$$u_{c*}k$$



# **Uncertainty Contributors**

### The minimum you should consider

#### **Reference Standard**

### Repeatability

### **Resolution, if you are calibrating an instrument**

From there it depends on the type of calibration as to what contributors should be considered. May include Specification or drift of equipment, Environmental effects "temperature, humidity, barometric pressure, Coefficient of Thermal Expansion, Physical Constant & Conversion Factors, Metrologist Effects



# **Uncertainty Contributors**

- Pipette Temperature/Humidity Effects
- Mass Air buoyancy associated with temperature, humidity and barometric pressure
- Electronic Drift associated with the Standard,
- Caliper, Micrometer, Gauge Blocks, Uncertainty associated with Thermal Expansion of material,
- Thermometer via liquid bath uniformity of bath
- IR Temperature influence of emissivity



# Uncertainty

#### 1) Type A,

Type A contributors to uncertainty are those that you have statistical data for. Use this data if you have it.

E.g., For **Standard Uncertainty** in the mean of repeated measurements (preferably 10 or more) use the experimental standard deviation of the mean.

#### 2) Type B

Type B contributors to uncertainty are those that you have no statistical data for; e.g.,

Manufacturer's specification

Professional judgement

Uncertainty in cal certificate for your reference standard



When using the uncertainty budget to estimate CMC for inclusion on its desired scope of accreditation, the calibration organization shall consider the performance of the "best existing device" available for each calibration sub-discipline. This means that for sources which can be expected to vary from calibration to calibration, identify the smallest contribution, which will occur when the conditions, which cause it, are at optimum and use these values in the estimate of CMC.

Remember you can never report an actual uncertainty which is less than the stated CMC on the scope of accreditation.  $\bigwedge^{?}$ 





# Sample Uncertainty Budget Pipette

👃 Pipette.pdf - Adob	e Acrobat	Reader DC (32-bit)										- [ [	ar 🛛 🗙
File Edit View Sig	gn Winde	ow Help											
Home Tools		PL-3 Uncertainty.pdf	Pipette.pdf ×								?	Sig	n In
🖹 🕁 Թ		Q	1	1 / 2	▶ ⊖ ⊕	135% 🔹 📑	• •				0	$\bowtie$	20
	rnis un	centainty budget applies	to the following type of thea:	surement.	Fipelle Dispenses	on Analytical Data	nces					*	
	Applica	ble range of measureme	ent:		200 ul								Q
	Followi	ng calibration procedure	no. and rev.:									-	B
	All unce	ertainties are expressed i	in units of		ul							-	-
	Numbe	r of significant figures for	r reporting of expanded unce	rtainty:	2	Dat	e prepared:	2014-07-24					
	Uncertainty budget prepared by: John Doe												L9
					Unantainte								
	i	Con	nponent of Uncertaint	y	Uncertainty, U(xi)	Distribution	Divisor	Std (	Jnc, u(xi)				17
	<u>1</u>	Uncertainty of Balan	ce from Cal Report		0.053	Normal, 2s	2.00	0.0265	uL				- <b>B</b>
	<u>2</u>	Linearity of Balance			0.15	Rectangular	1.73	0.0866	uL				
•	<u>3</u>	Repeatability of Pipe	ette		0.0918	Normal, 1s	1.00	0.0918	uL			•	4 ×n.
	<u>4</u>	Resolution of Balance	ce		0.01	Rect x 2	3.46	0.00289	uL				iU
	<u>5</u>	Temperature of Wat	er		0.00001	Rectangular	1.73	0.00000577	uL				
	<u>6</u>	Pipette Resolution			0.2	Rect x 2	3.46	0.057735027	uL				산
	<u>7</u>	Cubic Expansion Co	pefficient		0.00009	Rectangular	1.73	5.19615E-05	uL				
	<u>8</u>	Evaporation			0.001	Rectangular	1.73	0.00057735	uL				1
	<u>9</u>												
	<u>10</u>												
					com	bined standard un	certainty, u <sub>c</sub>	0.141	uL				
						covera	age factor, k	2	2				0
						expanded und	certainty, U <sub>c</sub>	0.282	2 uL				Cre
				Expan	ded uncertainty roun	ded UP to 2 signifi	cant figures	0.29	uL				$\sim$
													1.
												-	l⇒



## Power Torque Tools Calibration Uncertainty Budget

Copy of Power	er Tool 1	orque.pdf - Adobe Acrobat Read	er DC (32-l	bit)													7 X
File Edit View	Sign	Window Help															
Home Tools PL-3 Uncertainty.pdf			N	IISTIR 6919-	scale	Thermometer in liq			Copy of Power	Tool ×					?	Sig	n In
₿ ☆	ጭ	<b>a</b> 🔍		Ċ	D 🕁	1 / 2	<b>N</b> (	∋ ⊕	105% -	₽					0	$\bowtie$	20
		Description of Uncertainty Contributor	Type A/B	Parameter Uncertainty Limits	Parameter Units	Distribution	Level of Confidence	Coverage Factor (k)	Std Dev in Parameter Units	Effective Degrees of Freedom	Sensitivity Coefficient	Computed Uncertainty in Analysis Units	Effective Uncertainty in Analysis Units	Relative % of Total		*	0
	Α	MFG Spec.	В	1.4	ozf-in	Rectangular	100.00%	1.732	0.808290377	50	1	0.808290377	0.808290377	11.43			Pa
	В	Calibration Error	B	0.4	ozf-in	Normal	95.45%	2.000	0.199999756	50	1	0.199999756	0.199999756	2.83			ĿO
	С	Calibration Resolution	B	0.01	lbf-in	Resolution	100.00%	3.464	0.002886751	50	1	0.002886751	0.002886751	0.04			
	D	UUT Resolution	B	0.5	ozf-in	Resolution	100.00%	3.464	0.144337567	50	1	0.144337567	0.144337567	2.04			
		Operator Variable COS Error	В	0.4	ozt-in	U-Snaped	100.00%	1.414	0.282842712	50	1	0.282842712	0.282842712	4.00			
	F	DRP and Stability	D A	4 470529500	021-IN	Std Dov	100.00%	1.732	1.104700038	20	1	1.104/00038	1.104700038	16.3			
	H H	Nort and Stability	^	4.473320333	021-111	Siu Dev	1 Stu Dev	-	4.473320333		-	4.475320355	4.475520555	63.34			6
	<b>⊢</b> ∺⊢																
	i i																
	ĸ																
	L L															=	
	M																
	N																<b>1</b>
	0																
•	Ρ															•	KD.
	Q																
	R																
	S																S.
	T																
											_						
	V																1
	vv																
	-										-						
	<u> </u>			Variable 1	Variable 2	Corre	lation Coeffic	ient from 1 (	0 to -1 0			Corr Values					
		$u_{1} = \sqrt{u_{1}^{2} + u_{2}^{2} \dots u_{n}^{2} + 2\rho_{1,2}u_{n}u_{n}}$	_			o of 1.2 =				"Percent of Total account via u	" takes correlations into use of the "Effective	0.0000E+00	Uncertainty	GUM Method			· ·
		$\sqrt{-1}$ $\sqrt{-1}$ $\sqrt{-1}$ $\sqrt{-1}$	2			ρ of 1,2 =				Uncert	ainty" values.	0.0000E+00	Toolbox	Calculator			
		Bolded Percent of Total Values an	e > or = to:	25%	of Total Uncert	ainty	Tot	al Eff. Degree	es of Freedom (u)=	20.76	Std Unc =	4.7110E+00	4.7110E+00	Std Unc			Che
	User	$r = \sqrt{u_1^2 + \left( \left( \frac{u_1^2}{u_1^2 + u_2^2} \right) 2\rho_{1,2} u_1 u_2 \right)}$	95.45%	Student's t- Distribution	To achiev	e the stated "Leve	l of Confidence Degrees of Fre	using t-stat	istics and Effective alculated k-factor =	2.133	C Expanded Unc =	1.0049E+01	ozf-in	$v = \frac{u_T^{\Box^4}}{v_T}$			~
	w	$\sqrt{(u_1^* + u_2^*)}$	Level of Confidence	Normal Distribution	To ac	chieve the stated '	'Level of Confid distri	lence" for an ibution, the ca	assumed "Normal" alculated k-factor =	2.000	<ul> <li>Expanded</li> <li>Unc =</li> </ul>	9.4220E+00	ozf-in	$\sum_{l=1}^{n} \frac{u_l^{\Box^4}}{v_l}$			
																-	



## Electrical Simulation Temperature Calibration Uncertainty Budget

Samplek .	Type TC Sou	rce LINC simulation lerny odf - Ad	ohe Acroh	at Reader DC (	32-hit)												1 52
File Edit \	liew Sign	Window Help	obe Aciob	at header be (	JE DIG												
	view sign	Window Thep													-		
Home	Tools	SampleK Type TC S 3	ĸ												?	Sign	In
	-			-							_						
	7 ጥ			1		1 / 2	• (-	) ⊕	105% 🔻	Ē .	*				Ĉ	$\bowtie$	20
																	_
		Description of Uncertainty Contributor	Type A/B	Parameter Uncertainty	Parameter Units	Distribution	Level of Confidence	Coverage Factor (k)	Std Dev in Parameter Units	Effective Degrees of Erreedom	Sensitivity Coefficient	Computed Uncertainty in Analysis Units	Effective Uncertainty in Analysis Units	Relative % of Total		Î	Q
		744 Acc spop Mossuro		0.2	• •	Postangular	100.00%	1 7 9 9	0.172005081	F0	1	0.172005081	0.172005081	20.07			
	R	744 Acc spec measure 744 Cal LINC	B	0.08	°C	Normal	95.45%	2.000	0.173205061	50	1	0.039999951	0.039999951	0.21			Lo
	C	744 resolution	B	0.1	۰Č	Resolution	100.00%	3,464	0.028867513	50	1	0.028867513	0.028867513	6.64			
	D	744 CJ error	В	0.2	°C	Rectangular	100.00%	1.732	0.115470054	50	1	0.115470054	0.115470054	26.58			
	E	EMF Voltage error	Α	0.02	°C	U-Shaped	100.00%	1.414	0.014142136	50	1	0.014142136	0.014142136	3.26			
	F	DUT resolution	B	0.1	°C	Resolution	100.00%	3.464	0.028867513	50	1	0.028867513	0.028867513	6.64			
	G	Repetabality Data study	A	0.03391165	°C	Std Dev	1 Std Dev	1	0.03391165	29	1	0.03391165	0.03391165	7.81		_	Pa
	н																-0
														<u> </u>			_
	ĸ																
	L															=	
	M																
	N																<b>-</b>
	0																
•	P															•	.≍⊓
	Q																:U
	n S																
	Ť																224
	U																-
	V																-
	W																
	X																
	Y			Variable 1	Variable 0	Corro	lation Cooffici	ont from 1.0	to 1.0		l	Corr Voluce	l				
		$y = \int y^2 + y^2 + y^2 + 20$		Valiable I	Valiable 2	0.0f12=	auon coemic		0-1.0	"Percent of Total" ta account via use	kes correlations into of the "Effective	0.00005+00	Uncertainty	GUM Method			
		$u_x = \sqrt{u_1 + u_2 + \dots + u_n + 2p_{1,2}u_1}u_n$	2			0 of 1.2 =				Uncertain	ty* values.	0.0000E+00	Toolbox	Calculator			
		Bolded Percent of Total Values ar	e > or = to	25%	of Total Uncerta	ainty	Tota	Eff. Degrees	of Freedom (u)=	106.05	Std Unc =	2.1897E-01	2.1897E-01	Std Unc			1.
	ule	$ff = \left[u_1^2 + \left(\frac{u_1^2}{u_1^2 + u_2^2}\right)2\rho_{1,2}u_1u_2\right]$	95.45%	Student's t- Distribution	To achieve t	he stated "Level C	of Confidence" Degrees of Free	using t-statis dom, the calc	tics and Effective culated k-factor =	2.024	C Expanded Unc -	4.4318E-01	• C	$y = \frac{u_T^{\square^4}}{\square^4}$			~
	v	Vhere u <sub>1</sub> & u <sub>2</sub> are correlated	Level of Confidence	Normal Distribution	vormal To achieve the stated "Level of Confidence" for an assumed "Normal" $distribution$ distribution, the calculated k-factor = 2.000 $\odot$ Expanded 4.3795E-01 $\circ$ C												
																-	-2



# Significant Digits

As entered on the scope and uncertainty as reported on the calibration certificate, test report, or reference material certificate shall be expressed using no more than 2 significant digits and no insignificant digits. For guidance on methods to identify significant and insignificant digits as well as rules for rounding of numbers used to express the CMC or uncertainty refer to PJLA PL-4.

```
0.149 mg = 0.15 mg
2.85mg = 2.9 mg
1.16 lb = 1.2 lb
3.68 lb = 3.7 lb
```



The uncertainty covered by the CMC shall be expressed as the expanded uncertainty having a coverage probability of approximately 95 %. The unit of the uncertainty shall always be the same as the measured or in a term relative to the measured, e.g., percent,  $\mu$ V/V **Because of the ambiguity of definitions, the use of terms."PPM" and "PPB" are not acceptable.** 

PL-3 addresses in-house calibrations. Requirements concerning uncertainty budgets are the same as accredited calibration laboratories.



# Testing Labs

**7.6.3** A laboratory performing testing shall evaluate measurement uncertainty. Where the test method precludes rigorous evaluation of measurement uncertainty, an estimation shall be made based on an understanding of the theoretical principles or practical experience of the performance of the method.

When rigorous, mathematically, and statistically valid estimate of the measurement uncertainty may not be possible, so the requirements in ISO/IEC 17025:2017 7.6.3 would apply. In such cases the organization must identify all the components of uncertainty and make a "reasonable estimation". The "reasonable estimation" is to be based on knowledge of the performance of the method and on the measurement.



# Testing Labs

In those cases where a well recognized test method specifies limits to the values of the major sources of uncertainty of measurement and specifies the form of presentation of calculated results, the organization is considered to have satisfied ISO/IEC 17025:2017 clause 7.6.3 or ISO 15189:2012, Section 5.5.1.3. by following the test method and reporting instructions. Examples include ASTM, AOAC, BAM, USP, FDA, EPA, etc. methods as well as regulatory, legal methods – US CFR, EU/EC methods and associated reporting.



# Testing Labs

Other example include rapid method kits that specify limits to the values of the major sources (contributors) of uncertainty, as well as well-recognized rapid methods where kits are used to determine qualitative results, (for example, a semi-quantitative kit assay that reports qualitative results such as "presence" or "absence" based on a numeric value).

As a general rule qualitative or semi--quantitative tests or test that require personal judgement will not incorporate uncertainties.



Upon achieving accreditation, the uncertainty budgets and the decisions regarding sources of uncertainty shall be periodically reviewed and updated by the organization to reflect changes in the organization, its equipment, procedures or personnel that might influence the ability of the organization to perform specific calibrations or tests for which they are accredited. These changes shall be documented.

In other words, preparing your uncertainty budget would need to be reevaluated. An organization should always verify the uncertainty associated with standards used to detect any changes. Changes in facility and environment can impact uncertainty by affecting repeatability associated with the measurements.



*PJLA Policy on Measurement Uncertainty "PL-3"* ISO/IEC 17025:2017 (clause 7.8.4.1 a) requires calibration certificates to report the measurement uncertainty of the measurement result presented in the same unit as that of the measured or in a term relative to the measured (i.e., percent). Any deviation from this requirement would need to come under the realm of simplified reporting as specified in ISO/IEC 17025:2017(clause 7.8.1.3). This is only permissible, if agreed to by the customer during the contract review process. This agreement shall be documented

### From ISO/IEC 17025:2017

**7.8.4.1** In addition to the requirements listed in <u>7.8.2</u>, calibration certificates **shall** include the following:

a) the measurement uncertainty of the measurement result presented in the same unit as that of the measurand or in a term relative to the measurand (e.g. percent



# Decision Rule

Also, during contract review, the laboratory is required as per ISO/IEC 17025:2017 (clause 7.1.3) to define and capture the customer agreement as to the decision rule which will be employed when making statements of compliance. The decision rule is defined as a rule that describes how measurement uncertainty is accounted for when stating conformity with a specified requirement. If a statement of compliance is being

made on calibration reports, the agreed to decision rule is required as per ISO/IEC 17025:2017 (clause 7.8.6.1) to document the calibration report

For guidance in determining and selecting appropriate decision rules to meet these requirements specified in ISO/IEC 17025:2017, PJLA encourages the usage of ILAC G 8 "Guidance on Decision Rules and Statements of Conformity.



## **Decision Rule**



U = 95% expanded measurement uncertainty



# ILAC G8

#### https://ilac.org/publications-and-resources/ilac-guidance-series

#### Or google ILAC G documents





# PL-3 other revisions

Modified to include appropriate requirements and references for RMPs

Additional measurement uncertainty source for medical labs ISO/TS 20914:2019 added

See:

Section 7.0 REFERENCE MATERIAL PRODUCERS (RMPS) AND CERTIFIED REFERENCE MATERIAL PRODUCERS (CRMS)

Section 8.0 MEDICAL/CLINICAL LABORATORIES (15189)



#### PJLA Policy on Measurement Traceability "PL-2"



This time is allocated for answering questions. You should have a space provided for submitting questions.

Please keep questions related to the topic covered in this webinar;





# Save the Date

#### Next PJLA Webinar



#### Monday, Aug 30th 2021

ISO/IEC 17025:2017 Requirements for Corrective Action

